REMOVABLE "Y" LUER LOCK ADAPTOR
INSTRUCTIONS FOR USE



INTENDED USE: To provide dual access to a single lumen hemodialysis catheter.

INSTALLATION PROCEDURE:

1. Install "Y" Adaptor male luer onto female luer of catheter.

2. Turn base of adaptor clockwise until secure.

3. Flush adaptor with saline to ensure connection is secure.

4. Connect to bloodlines to begin treatment.

WARNINGS & PRECAUTIONS:

- Examine connection before each treatment to prevent catheter leakage.

- To prevent accidents, assure the security of all caps and bloodline connections prior to and between treatments.

- Discard biohazard according to facility protocol.

- Examine catheter female luer after each treatment for damage.

- Federal (USA) law restricts this device to sale by or on the order of a physician.

- Product is sterile in unopened, undamaged package. STERILIZED BY ETHYLENE OXIDE STERILE EO

- Single use only. DO NOT RE-USE. Re-Use may lead to infection or illness/injury.
- Do not use if package is damaged.
- Do not resterilize.

SYMBOL TABLE

5.1.1	Manufacturer*
5.3.4	Keep Dry*
5.6.3	Non-pyrogenic*
5.3.2	Keep Away from Sunlight*
5.2.3 STERILE EO	Sterilized Using Ethylene Oxide*
5.2.8	Do Not Use if Package is Damaged*
5.1.4	Use-by Date*
5.2.6	Do Not Resterilize*
5.1.5 LOT	Batch/Lot Number*
5.1.6 REF	Catalogue Number *
5.4.4	Caution, consult Accompanying Documents*
5.4.2	Do Not Re-use *
5.1.2 EC REP	Authorized Representative in the European Community
Rx Only	Prescription Use Only ***

***This symbol is in accordance with ISO 15223-1.**

***** FDA guidance Use of Symbols in Labeling.**

WARRANTY

Medcomp® WARRANTS THAT THIS PRODUCT WAS MANUFACTURED ACCORDING TO APPLICABLE STANDARDS AND SPECIFICATIONS. PATIENT CONDITION, CLINICAL TREATMENT, AND PRODUCT MAINTENANCE MAY EFFECT THE PERFORMANCE OF THIS PRODUCT. USE OF THIS PRODUCT SHOULD BE IN ACCORDANCE WITH THE INSTRUCTIONS PROVIDED AND AS DIRECTED BY THE PRESCRIBING PHYSICIAN.

Because of continuing product improvement, prices, specifications, and model availability are subject to change without notice. Medcomp® reserves the right to modify its products or contents without notice.

Medcomp® is a registered trademark of Medical Components, Inc.

Medical Components, Inc.
1499 Delp Drive
Harleysville, PA 19438 U.S.A.
Tel:215-256-4201
Fax:215-256-1787
www.medcompnet.com

EC REP
MPS Medical Product Service GmbH
Borngasse 20
35619 Braunfels
Germany

CE 0086