

ACCESSORIES IMPORTANT RISK INFORMATION

MICRO-STICK®

Indications for Use: The Micro-Stick® Set is indicated for percutaneous introduction of up to a 0.038 inch guidewire or catheter into the vascular system following a small 21 gauge needle stick. The Micro-Stick® Set is not intended for use in the coronary or cerebral vasculature..

PEELABLE INTRODUCERS

Indications for Use: The Valved Peelable Introducers are intended to obtain central venous access to facilitate catheter insertion into the central venous system.

The Super Sheath Introducer is intended to obtain central venous access to facilitate catheter insertion or placing pacing leads into the central venous system. The Super Sheath is compatible with a 0.018" or smaller guidewire.

Contraindications: The Valved Peelable Introducer and Super Sheath Introducer are not designed for use in the arterial system or as a hemostatic device.

DILATORS

Indications for Use: The Medcomp® vessel dilators are designed for percutaneous entry into a vessel in order to enlarge the opening of the vessel for the placement of a catheter in a vein or artery.

TESIO® REPAIR KIT

Indications for Use: The Medcomp® Tesio® Catheter is designed for Long Term Hemodialysis and Apheresis. It may be inserted percutaneously, and is ideally placed in the internal jugular vein. Although this catheter may be inserted into the subclavian vein, the internal jugular vein is the preferred site.

Contraindications: This extension is not intended for any use other than that which is indicated. To maintain peak performance of the extension, it is recommended that the extensions be replaced every 6 months.

SPLIT STREAM® REPAIR KIT

Indications for Use: : Repairs MedComp® Split Stream® Catheters

Contraindications: This catheter is intended for Long-Term vascular access only and should NOT be used for any purpose other than indicated in these instructions. To maintain peak performance of the Split-Stream® extension set, it is recommended that the extension set be replaced every 6 months

MEDCOMP® REPAIR KIT

Indications for Use: The MedComp® Repair Kit is indicated for use in replacing damaged female luer connectors, clamps, or repairing extensions where there is a minimum of 4.5cm viable extension tubing.
Repairs Medcomp® Catheters: Split Cath®

Contraindications: Do not use to repair catheters other than those specified above. Do not replace connector if tubing is swollen or displays signs of degradation. This repair kit should not be used for any purpose other than indicated in these instructions.

Refer to Instructions for Use provided with the product for complete instructions, warnings, precautions, and contraindications. Observe all instructions for use prior to using products. Failure to do so may result in patient complications.



TESIO® REPAIR KIT		
PBF11A	6.5F TESIO® CATHETER ARTERIAL EXTENSION SET	10/BOX
DBF11A	10F TESIO® CATHETER ARTERIAL EXTENSION SET	10/BOX

PBF11V	6.5F TESIO® CATHETER VENOUS EXTENSION SET	10/BOX
DBF11V	10F TESIO® CATHETER VENOUS EXTENSION SET	10/BOX

DBF11AVS	10F TESIO® VENOUS/ARTERIAL EXTENSION SETS	10/BOX
----------	---	--------

SPLIT CATH®, TITAN®, HEMO-FLOW® & HEMO-FLOW® XF MEDCOMP® REPAIR KIT		
DRPK-01	MEDCOMP® REPAIR KIT	5/BOX

SPLIT STREAM® REPAIR KIT		
DSSA	16F SPLIT STREAM® ARTERIAL EXTENSION SET	10/BOX
DSSV	16F SPLIT STREAM® VENOUS EXTENSION SET	10/BOX

TESIO® ARTERIAL EXTENSION SET

CONTENTS:

- (1) Arterial Extension Set
- (1) End Cap

TESIO® VENOUS EXTENSION SET

CONTENTS:

- (1) Venous Extension Set
- (1) End Cap

TESIO® VENOUS/ARTERIAL EXTENSION SET CONTENTS:

- (1) Arterial Extension Set
- (1) Venous Extension Set
- (2) End Cap

MEDCOMP® REPAIR KIT SET

CONTENTS:

- (1) Luer Assembly
- (1) Scissors
- (1) Adhesive Backed Drape
- (3) Clamps
- (2) End Caps

SPLIT STREAM® ARTERIAL EXTENSION SET CONTENTS:

- (1) Arterial Extension Set
- (2) Clamps
- (1) End Cap

SPLIT STREAM® VENOUS EXTENSION SET CONTENTS:

- (1) Venous Extension Set
- (2) Clamps
- (1) End Cap

Refer to the Table of Contents for Important Risk Information regarding this device.